

GelVue UV Transilluminator - Quick Guide


■ Description

1. GelVue UV Transilluminator is an ultra violet light box for use in gel imaging.

WARNING: UV radiation can be dangerous for unprotected eyes and skin, therefore, appropriate precautions must be taken. UV protective goggles or a face shield must be worn. Please contact our sales office for further details and accessories.

■ Operating instructions

1. Note the switch arrangement is as shown in the diagram below. Make sure that the On/Off switch is in the Off position and that the transilluminator is set to 100% (this can be adjusted if required).



2. Plug the mains input to the unit and supply. DO NOT switch the unit on before checking that safety precautions have been observed.
3. Now switch the unit on and ensure the switch lamp illuminates and the lamps in the unit illuminate evenly.
4. Please switch off the unit when not in use.
5. If unit fails to light up, check that the mains display is present, that the On/Off switch is in the On position. If no fault can be found please contact our service department for assistance.

CONTACT SYNGENE:
EUROPE:
Tel: +44 (0)1223 727123
Email: support@syngene.com

USA:
Tel: 800 686 4407/301 662 2863
Email: support@syngene.com

www.syngene.com
GVT0510